

الملخص الانجليزي للأبحاث المقدمة من الدكتورة / منى ابراهيم احمد
المدرس بقسم الامراض الصدرية و التدرن - كلية الطب - جامعة الفيوم
وذلك لتقديمها الى اللجنة العلمية الدائمة لأمراض الصدر توطئة للترقى لدرجة استاذ مساعد بالقسم

Paper 1:

Utility of GeneXpert MTB/RIF assay for the diagnosis of pulmonary and extra-pulmonary tuberculosis, A report from Egypt

Abstract:

Early diagnosis of tuberculosis continues to be a challenge for clinicians. The World Health Organization (WHO) guidelines recommend the application of GeneXpert MTB/RIF in extra-pulmonary tuberculosis (EPTB) diagnosis. This study aimed to test and compare the accuracy of the GeneXpert MTB/RIF assay to diagnose pulmonary tuberculosis (PTB) and EPTB, compared to bacterial culture and to composite reference standard (CRS). The GeneXpert assay diagnosed tuberculosis (TB) in 19.5 % of patients. With reference to bacterial culture, the sensitivity of this assay for detection of the pulmonary and extra-pulmonary specimens was perfect.

For pulmonary specimens, on using CRS; the detected sensitivity and specificity of the GeneXpert assay were 78.3 % and 99.1 %, respectively. However, for extra-pulmonary specimens, the sensitivity and specificity of the GeneXpert assay were 37.1 % and 99 %, respectively. In the current study, the GeneXpert assay showed almost perfect agreement with the bacterial culture for TB diagnosis. The diagnostic accuracy of the GeneXpert assay was high in ruling in, but not in ruling out of EPTB.

Keywords: GeneXpert MTB/RIF, Extra-pulmonary TB; Pulmonary TB, Egypt